

## Field-Cycling MRI – Techniques and Applications

*David J. Lurie<sup>1</sup>, Lionel Broche<sup>1</sup>, Changhoon Cho<sup>2</sup>, Gareth R. Davies<sup>1</sup>,  
Saadiya R. Ismail<sup>1</sup>, Dara Ó hÓgáin<sup>3</sup> and Kerrin J. Pine<sup>1</sup>*

<sup>1</sup>*Aberdeen Biomedical Imaging Centre, University of Aberdeen, AB25 2ZD, Scotland, UK*

<sup>2</sup>*Present address: MR Solutions Ltd., Merrow Business Park, Guildford, Surrey GU4 7WA, UK*

<sup>3</sup>*Present address: Stahlertherapie Abteilung, Landesklinikum Krems, Krems, Austria*

*<http://www.ffc-mri.org>*

In fast field-cycling (FFC) NMR, the magnetic field is switched between levels during the pulse sequence; it has been used for several decades for relaxometry measurements ( $T_1$  as a function of field strength) of small samples. The combination of FFC with magnetic resonance imaging (MRI) remains relatively uncommon, although its prevalence has been increasing in recent years [1].

One application of FFC is imaging free radicals using Field-Cycled Proton-Electron Double-Resonance Imaging (FC-PEDRI). This uses the Overhauser effect: irradiation of the free radical's ESR causes a transfer of polarisation from electron spins to coupled nuclear spins, resulting in a change in image intensity. Field-cycling allows the ESR irradiation to be applied at low field (hence relatively low frequency, and low non-resonant absorption), while NMR signal detection and imaging is carried out at higher field, to preserve SNR. We have constructed two FC-PEDRI scanners, both of which can also be used for FFC-MRI [2,3].

Relaxometric MRI is the imaging equivalent of field-cycling relaxometry. The aim is to obtain spatially-resolved  $T_1$ -dispersion data, by collecting images at a variety of evolution field strengths [1,4,5,6]. We have recently demonstrated methods for implementing relaxometry on localised regions defined from a pilot image [7]. We have also shown that FFC relaxometry can detect the formation of cross-linked fibrin protein from fibrinogen *in vitro*, in a model of the blood clotting process [1,8,9]. This relies on  $^{14}\text{N}$ - $^1\text{H}$  cross-relaxation phenomena, also known as “quadrupole dips” in the  $T_1$ -dispersion plot [10]. These reductions in  $T_1$ , occurring at Larmor frequencies equal to the  $^{14}\text{N}$  nuclear quadrupole resonances, reveal information about the concentration and conformation of immobilised protein molecules. In other recent work we have demonstrated that FFC-MRI can be used with tailored contrast agents which exhibit significantly different relaxivity over the range of field strengths accessible to an FFC-MRI scanner; in this way, the sensitivity of the experiment can be enhanced [11]. Another application of FFC is to study the phenomenon of magnetisation transfer contrast (MTC) as a function of magnetic field at low field [12,13].

- [1] Lurie D.J., Aime S., *et al.*, *Comptes Rendus Physique* **11**, 136-148 (2010).\*
- [2] Lurie D.J., Foster M.A., *et al.*, *Phys.Med.Biol.* **43**, 1877-1886 (1998).\*
- [3] Lurie D.J., Davies G.R., *et al.*, *Magn.Reson.Imaging* **23**, 175-181 (2005).\*
- [4] Carlson J.W., Goldhaber D.M., *et al.*, *Radiology* **184**, 635-639 (1992).
- [5] Lurie D.J., 1<sup>st</sup> Symposium on Field-Cycling NMR Relaxometry, Berlin, p5, (1998).\*
- [6] Ungersma S.E., Matter N.I., *et al.*, *Magn.Reson.Med.* **55**, 1362-1371 (2006).
- [7] Pine K.J., Davies G.R. and Lurie D.J., *Magn.Reson.Med.* **63**, 1698–1702 (2010).\*
- [8] Broche L.M., Ismail S. *et al.*, Proc. 18<sup>th</sup> ISMRM, Stockholm, Sweden, p915 (2010).\*
- [9] Broche L.M., Ismail S. *et al.*, *Magn.Reson.Med.*, In Press (2011).
- [10] Kimmich R., Winter F. *et al.*, *J.Magn.Reson.*, **68** (1986) 263-282.
- [11] Ó hÓgáin D., Davies G.R., *et al.*, *Phys. Med. Biol.* **56**, 105-115 (2011).\*
- [12] Choi C.-H., Davies G.R. and Lurie D.J., *J.Magn.Reson.* **204**, 145-149 (2011).\*
- [13] Choi C.-H., Hutchison J.M.S. and Lurie D.J., *J.Magn.Reson.* **207**, 134-139 (2011).\*

\*These references are available at <http://www.ffc-mri.org/publications.shtml>