

# Manganese Liposomes as Molecular Probes using Fast Field-Cycling MRI

Dara Ó hÓgáin<sup>1,\*</sup>, Simona Baroni<sup>2</sup>, Silvio Aime<sup>3</sup>, Gareth R. Davies<sup>1</sup> and David J. Lurie<sup>1</sup>

<sup>1</sup>Aberdeen Biomedical Imaging Centre, University of Aberdeen, Aberdeen, AB25 2ZD, UK;

<sup>2</sup>Invento S.r.l., Torino, Italy; <sup>3</sup>Department of Chemistry, University of Torino, Torino, Italy

\*e-mail: [r01do7@abdn.ac.uk](mailto:r01do7@abdn.ac.uk)

## Purpose:



Fast Field Cycling (FFC) MRI is able to switch between magnetic fields during the application of a pulse sequence. In this way it is potentially able to exploit the different  $T_1$  dispersion behaviour of tissue and contrast agents to optimise contrast between normal and pathological tissues. In particular, FFC-MRI allows access to a new contrast mechanism known as  $\Delta R_1$  imaging in which contrast depends on the change in  $R_1$  between different fields [1]. In this work, a proof-of-principle study has been carried out using contrast agents designed specifically for use with FFC-MRI which show large changes in  $R_1$  between different magnetic fields. These contrast agents are detected with much greater sensitivity in a  $\Delta R_1$  image compared with standard  $T_1$  weighted images.

## Materials and Methods:

A new contrast agent consisting of Mn[II] ions surrounded by POPC:PEG (95:5) liposomes diluted in a HEPES buffer solution was designed for use with FFC-MRI. The solution was first diluted to five different concentrations; A: 0.15 mM, B: 0.1 mM, C: 0.08 mM, D: 0.06 mM, and E: 0.04 mM respectively. Sample F containing 1.0 mM CuSO<sub>4</sub> was also prepared in de-ionised water in order to provide comparison with the Mn[II] contrast agent, as CuSO<sub>4</sub> unlike Mn[II] shows almost no change in  $R_1$  between field strengths. The samples were imaged using a home-built, whole-body FFC-MRI scanner with detection at  $B_0 = 58.7$  mT (2.499 MHz) [2].

## Results:

$T_1$  weighted images were acquired initially at 58.7 mT with evolution times of 10 ms, 100 ms, 220 ms, 330 ms, 440 ms and 600 ms (Figure 1A, images (i – vi) respectively). The images were then acquired at 2 mT using the same evolution times. Using these data, maps of  $R_1$  values were calculated using a program written in MATLAB (MathWorks, MA, USA), for each pixel in a set of images.  $R_1$  maps at 59 mT and 2 mT were then subtracted to give a  $\Delta R_1$  value for each pixel. Figure 1B shows a  $\Delta R_1$  map for each of the samples (A – F) in the imaging phantom. Figure 2 shows the ratio between the image intensities of the Mn[II] samples and the CuSO<sub>4</sub> sample, comparing the  $\Delta R_1$  values to  $T_1$  weighted images at 2 mT, and at 59 mT.



## Conclusions:

The FFC-MRI  $\Delta R_1$  mapping technique takes advantage of the  $R_1$  dispersion of the Mn[II] liposomal contrast agent, resulting in higher contrast between the Mn[II] and the 1.0 mM CuSO<sub>4</sub> samples compared with  $T_1$  weighted imaging. Typically normal tissue exhibits only a small change in  $R_1$  between 2.5 and 10 MHz [3], and as such,  $\Delta R_1$  mapping could be used to enhance contrast between regions containing contrast agent and surrounding areas of normal tissue.

## References:

- [1] Alford *et al.*, Magn. Reson. Med., **61**, 795 (2009).
- [2] Lurie *et al.*, Phys. Med. Biol., **43**, 1877 (1998).
- [3] Bottomley *et al.*, Med. Phys., **11**, 425. (1984).